

DECLARATION OF CONFORMITY

We the manufacturer:

493K Limited, 23 Watch Hill Road, Ballyclare BT39 9QW, United Kingdom.

declare under our sole responsibility that the product :-

K-PAQ Receiver Data Relay Model TXR-2A-M/S

to which this declaration relates is in conformity with the following standard(s) or other normative document(s) as set out in the following provisions:-**2014/30/EU (EMC)****EN 61326-1:2013 Class A Table 2**

Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements

EN 55011:2009+A1:2010 Class A Group 1

Industrial, scientific and medical equipment – Radio-frequency disturbance characteristics – Limits and methods of measurement

EN 61000-3-2:2014

Electromagnetic compatibility (EMC) -- Part 3-2: Limits - Limits for harmonic current emissions (equipment input current ≤ 16 A per phase)

EN 61000-3-3:2013

Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection

2014/35/EU (LVD)**EN 61010-1:2010**

Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements

EN 61010-2-030:2010

Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-030: Particular requirements for testing and measuring circuits

EN 61010-2-201:2013

Safety requirements for electrical equipment for measurement, control and laboratory use –Part 2-201: Particular requirements for control equipment

Place of issue: 493K Limited, 23 Watch Hill Road,
Ballyclare, Co Antrim. BT39 9QW, UK.
t: +44 (0)28 93359922 e: sales@493k.com

Date of issue: 21 December 2021

Signature:



Name: Dr Gareth W. G. McDowell

Position: Managing Director